

## 水中の低級アルコールの分析

## Analysis of Lower alcohols in water




Fig. Chromatogram of lower alcohols in water.

## 【Analytical conditions】

Column: **G-column G-205** 40 m L. × 1.2 mm I.D. 5 μm film thickness  
Flow rate: He 14.9 mL/min  
C.Temperature: 50°C  
Detection: FID  
Injection volume: 1 μL  
System: GC-9A(Shimadzu Co.)